



PAPER

OPEN ACCESS






RECEIVED
16 May 2021REVISED
2 July 2021ACCEPTED FOR PUBLICATION
21 July 2021PUBLISHED
1 September 2021

Original content from this work may be used under the terms of the [Creative Commons Attribution 4.0 licence](#).

Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.



Simulating NEMA characteristics of the modular total-body J-PET scanner—an economic total-body PET from plastic scintillators

P Moskal^{1,2,*} , P Kowalski³ , R Y Shopa³, L Raczyński³, J Baran⁴, N Chug^{1,2}, C Curceanu⁵, E Czerwiński^{1,2} , M Dadgar^{1,2}, K Dulski^{1,2}, A Gajos^{1,2}, B C Hiesmayr⁶ , K Kacprzak^{1,2}, Ł Kapłon^{1,2}, D Kisiielewska^{1,2}, K Klimaszewski³, P Kopka³, G Korcyl^{1,2}, N Krawczyk^{1,2}, W Krzemień⁷, E Kubicz^{1,2}, Sz Niedźwiecki^{1,2}, Sz Parzych^{1,2}, J Raj^{1,2}, S Sharma^{1,2}, S Shivani^{1,2}, E Stępień^{1,2} , F Tayefi^{1,2} and W Wiślicki³

¹ Faculty of Physics, Astronomy and Applied Computer Science, Jagiellonian University, 30-348 Cracow, Poland

² Total-Body Jagiellonian-PET Laboratory, Jagiellonian University, 30-348 Cracow, Poland

³ Department of Complex Systems, National Centre for Nuclear Research, 05-400 Otwock-Świerk, Poland

⁴ Institute of Nuclear Physics Polish Academy of Sciences, 31-342 Cracow, Poland

⁵ INFN, Laboratori Nazionali di Frascati, I-00044 Frascati, Italy

⁶ Faculty of Physics, University of Vienna, A-1090 Vienna, Austria

⁷ High Energy Physics Division, National Centre for Nuclear Research, 05-400 Otwock-Świerk, Poland

* Author to whom any correspondence should be addressed.

E-mail: p.moskal@uj.edu.pl

Keywords: J-PET, NEMA characteristics, large axial field-of-view, total-body PET, medical imaging, nuclear imaging

Abstract

The purpose of the presented research is estimation of the performance characteristics of the economic total-body Jagiellonian-PET system (TB-J-PET) constructed from plastic scintillators. The characteristics are estimated according to the NEMA NU-2-2018 standards utilizing the GATE package. The simulated detector consists of 24 modules, each built out of 32 plastic scintillator strips (each with cross section of 6 mm times 30 mm and length of 140 or 200 cm) arranged in two layers in regular 24-sided polygon circumscribing a circle with the diameter of 78.6 cm. For the TB-J-PET with an axial field-of-view (AFOV) of 200 cm, a spatial resolutions (SRs) of 3.7 mm (transversal) and 4.9 mm (axial) are achieved. The noise equivalent count rate (NECR) peak of 630 kcps is expected at 30 kBq cc⁻¹. Activity concentration and the sensitivity at the center amounts to 38 cps kBq⁻¹. The scatter fraction (SF) is estimated to 36.2 %. The values of SF and SR are comparable to those obtained for the state-of-the-art clinical PET scanners and the first total-body tomographs: uExplorer and PennPET. With respect to the standard PET systems with AFOV in the range from 16 to 26 cm, the TB-J-PET is characterized by an increase in NECR approximately by factor of 4 and by the increase of the whole-body sensitivity by factor of 12.6 to 38. The time-of-flight resolution for the TB-J-PET is expected to be at the level of CRT = 240 ps full width at half maximum. For the TB-J-PET with an AFOV of 140 cm, an image quality of the reconstructed images of a NEMA IEC phantom was presented with a contrast recovery coefficient and a background variability parameters. The increase of the whole-body sensitivity and NECR estimated for the TB-J-PET with respect to current commercial PET systems makes the TB-J-PET a promising cost-effective solution for the broad clinical applications of total-body PET scanners. TB-J-PET may constitute an economic alternative for the crystal TB-PET scanners, since plastic scintillators are much cheaper than BGO or LYSO crystals and axial arrangement of the strips significantly reduces the costs of readout electronics and SiPMs.

1. Introduction

Positron emission tomography (PET) is a well established diagnostic method enabling detection of a tissue pathology on a molecular level before it evolves to the functional or morphological abnormalities (McKenney-Drake *et al* 2018, Schmall *et al* 2019). Currently, routine PET imaging with devices of about 20 cm axial field-of-view (AFOV) (Grant *et al* 2016, Van Sluis *et al* 2019), in a single bed position, enables the diagnosis of individual

Table 1. Basic properties of LYSO crystal, BGO crystal and BC-408 (equivalent of EJ-200) plastic scintillators important for the design of the PET systems. The values of attenuation coefficients and the fractions of photoelectric effect were extracted from the data base maintained by the National Institute of Standards and Technology (National Institute of Standards and Technology 2020).

Scintillator	Density (g cm^{-3})	Light output (photons MeV^{-1})	Decay time (ns)	Fraction of photoelectric effect (%)	Light attenuation length (cm)	Linear absorption coefficient for 511 keV photons (cm^{-1})
LYSO	7.1–7.4 (Mao <i>et al</i> 2013)	33 200 (Crystals 2018a)	36 (Crystals 2018a)	31–32 (National Institute of Standards and Technology 2020)	21–40 (Crystals https://crystals.saint-gobain.com , Vilardi <i>et al</i> 2006, Mao <i>et al</i> 2008)	0.82–0.87 (National Institute of Standards and Technology 2020)
BGO	7.13 (Crystals 2016)	8000–10 000 (Crystals 2016)	300 (Crystals 2016)	41 (National Institute of Standards and Technology 2020)	55 (Chen <i>et al</i> 2004)	0.96 (National Institute of Standards and Technology 2020)
BC-408/ EJ-200	1.023 (Crystals https://crystals.saint-gobain.com)	10 000 (Crystals 2018b)	2.1 (Crystals https://crystals.saint-gobain.com)	$6.3 \cdot 10^{-5}$ (National Institute of Standards and Technology 2020)	380 (Crystals https://crystals.saint-gobain.com)	0.096 (National Institute of Standards and Technology 2020)


organs only, and the diagnosis of the whole-body requires a combination of series of sequential images obtained from many patient positions in the scanner, thus only a time-dependent scan of the total-body is available (Houshmand *et al* 2015). With the advent of the total-body PET (TB-PET), precision medicine will be enhanced with a new toolbox that allows for the simultaneous molecular imaging of the whole human body, providing concurrent imaging of metabolic rate in near and distant organs (Cherry *et al* 2017, 2018, Badawi *et al* 2019, Efthimiou 2020, Jones 2020, Karp *et al* 2020, Moskal and Stępień 2020, Surti *et al* 2020a, 2020b, Vandenberghe *et al* 2020, Zhang *et al* 2020). Thanks to the high sensitivity, TB-PET enables the extreme reduction of the whole-body imaging duration or the lessen of the radiopharmaceutical dose (Majewski 2020), thus opening perspectives for application of PET to the wider group of patients (e.g. children Nardo *et al* (2020)) or patients suffering from systemic diseases (Yamashita *et al* 2014, Nakajima *et al* 2017, McKenney-Drake *et al* 2018, Borja *et al* 2020, Vandenberghe *et al* 2020). By the introduction of uExplorer, the first total-body PET, to the clinical practice (Badawi *et al* 2019) it was demonstrated (Zhang *et al* 2019, 2020) that in addition to the static standardized uptake value images, TB-PET may also deliver a kinetic model based parametric imaging of all tissues in the body, simultaneously. These new capabilities open promising prospects for quantitative improvements of diagnostic and prognostic assessments of e.g. oncological, cardiological and neurological diseases (Majewski 2020).

Yet, the high costs of the TB-PET scanner based on LYSO crystal scintillators, estimated to around \$ 10 million or more (Cherry *et al* 2018), is a serious obstacle to the widespread use of this modality in clinical practice, including the medical research clinics. Therefore, reducing the cost of TB-PET production has become one of the important challenges taken by many research groups, and constitutes one of the hot research topics in this field. The considered solutions include: reduction of scintillator thickness (Surti *et al* 2013, 2020a), reduction of number of detectors by arranging them in sparse configuration (Zhang *et al* 2019, Zein *et al* 2020, Zhang and Wong 2017) or taking advantage of the Cherenkov light to improve timing properties with BGO crystals (Brunner and Schaart 2017, Gonzalez-Montoro *et al* 2017, Zhang and Wong 2017, Cates and Levin 2019, Kwon *et al* 2019, Gundacker *et al* 2020). In order to take advantage of the Compton scattering there is also an ongoing research aiming at combining standard PET tomography with Compton cameras (Grignon *et al* 2007, Donnard *et al* 2012, Lang *et al* 2012, Oger *et al* 2012, Lang *et al* 2014, Thirolf *et al* 2015, Hamidreza *et al* 2017, Aya *et al* 2017, Kuramoto *et al* 2017, Kenji *et al* 2020, Mizuki *et al* 2020, Yoshida *et al* 2020).

LYSO crystals account for about 50% of the total costs of the TB-PET scanner, while the rest of the costs comes mainly from readout electronics and SiPMs. An application of BGO crystals may reduce costs of the scintillators only by a factor of about 2 (Vandenberghe *et al* 2020). Thus, as it was argued in reference (Moskal and Stępień 2020), that the reduction of crystal thickness or exchange of the LYSO by BGO crystals will not lead to a sufficiently significant reduction of production costs. Therefore, one of the promising, economic solution for the construction of the TB-PET scanner is the exchange of the expensive crystals to cost-effective plastic scintillator strips arranged axially, further on referred to as Jagiellonian PET (J-PET) (Moskal *et al* 2011, 2014, 2015, 2016a, Niedźwiecki *et al* 2017, Moskal and Stępień 2020, Sharma *et al* 2020b, 2020a). The costs of components for the total-body plastic J-PET are expected to be about 5 times less than the crystal-based total-body PET (Moskal and Stępień 2020).

Table 1 compares basic properties of plastic scintillator with BGO and LYSO crystals which are currently used in PET detectors. The much higher density of crystals with respect to plastics, and the more effective registration of annihilation photons, can be compensated by the multi-layer geometry (Moskal *et al* 2016a) possible with the axial arrangement of plastic strips with the readout at the ends (Moskal *et al* 2014). The significantly lower light attenuation in plastic with respect to crystals (e.g. plastic scintillator BC-480 is characterized by 7 and 18 times longer light attenuation length compared to BGO and LYSO crystals, respectively) enables the effective light transport even up to 2 m long plastic strips. Moreover the negligible fraction of photo-electric effect for the interaction of 511 keV photons in plastic scintillators does not preclude the possibility of the scatter fraction (SF) reduction. When using plastic scintillators the scattering in the patient may be suppressed based on the measurement of the energy deposition due to the Compton interaction (Moskal *et al* 2014, 2016a).

Total-body J-PET (TB-J-PET) may constitute an economic alternative for the crystal TB-PET scanners, since plastic scintillators are more than an order of magnitude less expensive than BGO crystals, and plastic PET with axially arranged scintillator strips reduces also significantly costs of readout electronics and SiPMs. The reduction of the electronics cost may be achieved due to the fact that the readout (except the wavelength shifters (WLS) strips) is placed at the ends of the cylindrical detector compared to the coverage of the cylinder surface in case of the radially arranged blocks of crystal PET detectors (Moskal and Stępień 2020). Prospects and clinical perspectives of TB-J-PET imaging using plastic scintillators were recently described in Moskal *et al* (2019a), (2019b), Moskal and Stępień (2020). Prospects for fundamental physical questions can be found in Kamińska *et al* (2016), Moskal *et al* (2016b), (2018), Hiesmayr and Moskal (2019), Gajos (2020). In this article we assess the performance characteristics of TB-J-PET constructed from plastic scintillator strips. The spatial resolution (SR),



sensitivity (S), SF, noise equivalent count rate (NECR) and image quality (IQ) for the TB-J-PET are estimated according to the National Electrical Manufacturers Association NEMA NU 2-2018 standards (NEMA 2018) by using Geant4 Application for Tomographic Emission (GATE) (Jan *et al* 2004, 2011, Sarrut *et al* 2014, 2021) and the dedicated analysis software developed by the J-PET group (Kowalski *et al* 2018, Krzemień *et al* 2020). We assess NEMA characteristics for few TB-J-PET configurations. Transaxial field-of-view with diameter of 78.6 cm is assumed for all studied geometries, while AFOV of 200 cm (total-body) and 140 cm (head and torso) is considered. In addition, the influence on the result due to the unknown depth-of-interaction (DOI) is studied comparing results obtained assuming: (i) an ideal case that the true interaction point is known, (ii) a case when the DOI is not known, and (iii) assuming the resolution for the determination of DOI (full width at half maximum (FWHM)) is equal to $\text{FWHM}(\text{DOI}) = 10$ mm. Moreover studies of SR as a function of maximum allowed axial distance between the interaction points (the maximum accepted oblique angle of the line-of-response (LOR)) were also conducted.

In the section Materials and methods (section 2) we define the geometry and structure of the TB-J-PET scanner, properties of the materials constituting the detector and the assumed temporal and SRs. This section describes also briefly the methods of how the NEMA characteristics are evaluated. Next, results of simulations and analysis are presented in the section Results (section 3), which is followed by the Discussion (section 4) section including comparison of TB-J-PET characteristics to the performance characteristics of current clinical solutions and the total-body crystal-based PET scanners, in particular to the total-body uExplorer (Cherry *et al* 2017, Zhang *et al* 2017, Cherry *et al* 2018, Badawi *et al* 2019) and PennPET Explorer (Karp *et al* 2020) systems. While the NEMA norms were designed for PET systems with small AFOV, they may not be adequate for the total-body PET scanners with AFOV exceeding 70 cm long phantoms and emission sources required by the NEMA NU 2-2018 (NEMA 2018). Therefore, in the Discussion section the results obtained according to the NEMA norm are compared to the values obtained for the longer sources with lengths of 140 and 200 cm.

2. Materials and methods

Plastic scintillators ($1.021.06 \text{ g cm}^{-3}$) (Eljen Technology https://eljentechnology.com/images/technical_library/Physical_Constants_Plastic.pdf, Crystals <https://crystals.saint-gobain.com>) are about 7 times less dense than LYSO crystals ($7.07.4 \text{ g cm}^{-3}$) (Mao *et al* 2013) and the linear attenuation coefficient for the 511 keV photons is much higher for LYSO and BGO ($\mu = 0.83 \text{ cm}^{-1}$ and $\mu = 0.96 \text{ cm}^{-1}$, respectively) than for plastic scintillator ($\mu = 0.096 \text{ cm}^{-1}$) (see table 1). This implies that in a single scintillator layer with thickness of 2–3 cm




Figure 2. CRT values as a function of the AFOV expected while using BC-408 scintillator. Values obtained for the radial thickness of 1.5 cm and 3.0 cm, in the case of unknown DOI, are indicated by a solid line with dots and a dotted line, respectively. The result for ideal case with a known DOI is shown by a solid line.

(typical thickness of scintillators used in PET (Conti 2009, Vandenberghe *et al* 2020)) most of annihilation photons hitting the detector are interacting in case of LYSO and BGO (81%–92% and 85%–94%, respectively) but only about a quarter (17%–25%) in case of plastic scintillators. Therefore, in case of TB-J-PET the registration efficiency is increased significantly by the application of multi-layer detection system with an effective total thickness larger than 3 cm in the case of two layers, see figure 1, e.g. for 2–3 cm thick layers the registration efficiency would be about 44%.

2.1. Detector system configuration

In this article we consider a design of TB-J-PET with a double layer geometry as it is presented in the upper panel of figure 1. The detector consists of 24 modules, each including 32 scintillator strips arranged in two layers with the additional 3 mm thick layer of WLS (lower panel of figure 1). WLS layer is used for the reconstruction of the axial coordinate of the annihilation photon's interaction point (Smyrski *et al* 2014, 2017, Shivani *et al* 2020). Scintillator strips are rectangular in cross section with dimensions of 0.6 cm \times 3 cm. In this article, two cases are considered: strips with the length of $L = 140$ cm and $L = 200$ cm. The modules are arranged in regular 24-sided polygon circumscribing a circle with the diameter of $D = 78.6$ cm.

The diameter and length of simulated geometries were set to be close to the one chosen for the total-body scanners: uExplorer ($L = 194$ cm, $D = 78.6$ cm) (Badawi *et al* 2019, Spencer *et al* 2020) and PennPET Explorer ($L = 140$ cm, $D = 81$ cm) (Karp *et al* 2020).

2.2. Spatial and temporal resolution

In order to take into account the detector spatial and temporal resolution the simulated time and axial position of photons' interactions were smeared according to a Gaussian distribution. In case of axial resolution the FWHM of 5 mm of the Gaussian was assumed. Such a resolution is expected when applying a layer of WLS strips arranged perpendicularly to the scintillator strips (Smyrski *et al* 2014, 2017, Kowalski *et al* 2018).

For the estimation of the influence of the uncertainty of DOI reconstruction three scenarios are considered: (i) ideal case with DOI known from the simulations, (ii) standard case when DOI is estimated as center of the scintillator in the transaxial cross section, and (iii) assuming that the resolution of the determination of DOI can be approximated by the Gaussian function with a FWHM = 10 mm.

The expected values of coincidence resolving time (CRT) expressed as FWHM were estimated using a simulation method described in Moskal *et al* (2016a) and assuming that detector is built from BC-408 scintillators with a photon absorption length of 380 cm (table 1). Figure 2 presents the dependence of expected CRT value as a function of the AFOV for the case of unknown DOI in scintillators with thickness of 1.5 cm (solid line with dots), 3 cm (dotted line) and for the ideal case with known DOI (solid line). For the estimation presented in this article CRT values obtained for the 3 cm thickness with unknown DOI were used, which represents the worst case scenario.




Figure 3. Images of the point source placed in position of (1, 0, 0) cm (2 left columns) and in position of (20, 0, 0) cm (2 right columns) reconstructed with two methods: FBP3DRP and TOF-FBP. Results for AFOV of 140 cm are shown.

In order to account for the time resolution of the detector system, a simulated time of annihilation photon's interaction was smeared according to the Gaussian distribution corresponding to $CRT = 190$ ps and $CRT = 240$ ps for AFOV of 140 cm and 200 cm, respectively.

2.3. Simulation tools

Calculations are preformed using the GATE software (Jan *et al* 2004, 2011, Sarrut *et al* 2014, 2021) which enables generation of back-to-back annihilation photons of the sources as defined by the NEMA norm, and simulation of interaction of these photons in the detector material. Using the dedicated analysis programs the list-mode set of coincidences is obtained (Kowalski *et al* 2015, 2016, 2018). Finally, simulated times and positions of interactions are smeared according to resolutions discussed in the previous sub-section. Further on in order to reduce contribution of events with photons scattered in the phantom and in the detector, the simulated events are filtered using criteria based on the correlations between the hit time, hit position and energy deposition of annihilation photons in the detector. The event selection method is described in detail in Kowalski *et al* (2018).

2.4. Image reconstruction method

3D bin-mode filter-back-projection (FBP) with re-projection (FBP3DRP), implemented in Software for tomographic image reconstruction (STIR) (Thielemans *et al* 2012, Khateri *et al* 2019), was employed in our earlier works (Shopa *et al* 2017, Kowalski *et al* 2018, Moskal *et al* 2020). It is, however, impractical for 2-layer 2 m long J-PET due to the limitations of the geometries that the STIR framework supports: both multi-layer and continuous scintillators cannot be defined. Therefore, as a temporary solution, positions of scatterings inside scintillator strips (hits) are remapped onto a virtual single cylindrical layer in transverse plane and onto discrete rings along axial direction. Such procedure, though, may impose additional distortion and worsen axial resolution, undermining the benefits of re-projection at the same time (see the discussion in Shopa *et al* (2017)).

Alternatively, we developed a time-of-flight (TOF) FBP algorithm (TOF FBP), based on the idea presented in Conti *et al* (2005). Projection data is weighted twice: with a filter in frequency space and during back projection for each TOF bin. One could also define a more general, non-bin definition based on the summation over LORs. The reconstructed image f for an arbitrary voxel \mathbf{v} is estimated using back-propagation operator \mathcal{B} as:

$$f(\mathbf{v}) = \sum_{i=1}^{N_{\text{LOR}}} f_i(\mathbf{v}) = \sum_{i=1}^{N_{\text{LOR}}} \mathcal{B} p_i^F, \quad (1)$$

where each filtered projection element

$$p_i^F \equiv p_i^F(s, \phi, \zeta, \theta, t) = \mathcal{F}^{-1}\{W(\nu_s)\mathcal{F}[p_i(s, \phi, \zeta, \theta)]\} \cdot h(t - t_i). \quad (2)$$

s, ϕ, θ, ζ are coordinates in projection domain of sinograms p_i (Bailey *et al* 2005), \mathcal{F} and \mathcal{F}^{-1} are 1D Fourier and inverse Fourier transform operators, respectively, which serve to operate between space domain (projection coordinate s) and frequency domain (coordinate ν_s). $W(\nu_s)$ is the ramp filter, i denotes the index of LOR, $h(t)$ is TOF kernel and t_i is the temporal parameter which defines the position of annihilation along LOR. For each (i th)

coincidence, the dimensionality of p_i^F is reduced from \mathbb{R}^5 to \mathbb{R}^2 , since only s and t are not fixed, i.e.

$$p_i^F \equiv p_i^F(s, \phi_i, \zeta_i, \theta_i, t).$$

The sum in equation (1) requires multiple integration over every LOR, which could be time consuming. Besides, TOF normalisation is essential, because each back-projection $\mathcal{B}p_i^F$ is not made over the entire image plane, but is restricted by $h(t)$ kernel instead. However, this can be ignored for the case of 1 mm NEMA source, since it is definitely smaller than the minimal sampling for the displacement coordinate s . Furthermore, formula (1) can be redefined directly in image space as asymmetrical three-component kernel: one along LOR, which represents TOF function $h(t)$, second—along Z -axis, reflecting the uncertainty of hit position, and the third—as image-domain ramp filter, which operates on transverse plane along the direction, orthogonal to LOR. In this work, we shall use Gaussian definition for TOF and Z components, as it proved to be consistent with FBP3DRP and was successfully tested on both multi-layer and 2 m long scanners (Moskal *et al* 2020, Shopa 2020).

Each resulting image was additionally corrected by the sensitivity map of the TB-J-PET, generated beforehand in GATE using Monte Carlo method. For comparison, selected data for 140 cm long scanner were reconstructed using FBP3DRP. Since STIR operates with single layer geometries only, in the latter case hits were projected on ideal single layer cylindrical scanner. Furthermore, each strip was axially divided into 256 discrete sections—maximal value allowed for 1-byte representation of axial coordinate, set in SAFIR (Becker *et al* 2017)—a STIR extension which we employed for the conversion of data into suitable Interfile format. Each voxel of the reconstructed image for the STIR FBP3DRP had dimensions of 1.50 mm along x and y axes and 2.73 mm along the z axis. For our TOF FBP implementation a voxel had dimensions of 1.56 mm along x and y axes and 2.59 mm along the z axis.

Comparison of reconstruction results obtained using FBP3DRP and TOF FBP, for 140 cm long tomograph and the point source placed in position of (1, 0, 0) cm is presented in figure 3.

The outcomes are fairly consistent with each other. Therefore, further on, for the estimation of SR we will use the TOF FBP which enables image reconstruction with multi-layer geometry without necessity of remapping interaction points onto single layer.

2.5. NEMA-NU2-2018 norm

2.5.1. Spatial resolution

The standard, approved for the estimation of SR by National Electrical Manufacturers Association (NEMA) (NEMA 2018), defines point-like source for the measurements and its positioning inside PET scanner: three transverse locations ($x = 1$ cm, 10 cm and 20 cm) and two axial—at the centre and at $3/8$ of AFOV. The SR (point spread function—PSF) is defined as a FWHM of the reconstructed image, calculated for each direction separately. For the image reconstruction the TOF FBP method was utilised as explained above in section 2.4.

2.5.2. SF and NECR


SF of the PET scanner quantifies the sensitivity of the detector to scattered radiation. It is expressed as a ratio between the rates of scattered coincidences and the sum of scattered and true coincidences: $SF = \frac{S}{S + T}$. It is measured (or simulated) for relatively low source activity, such that the contribution of accidental coincidences is negligible.

NECR is the characteristic that shows the influence of scattered and random coincidences on the performance of the scanner as a function of the source activity and it is a measure of the effective sensitivity of the scanner (Conti 2009). The NECR is defined as: $NECR = \frac{T^2}{T + S + R}$, where T stands for the rate of true coincidences, S —scattered coincidences, R —random (accidental) coincidences.

For both SF and NECR, simulated phantom is a solid cylinder made of polyethylene with an outside diameter equal to 20.3 cm and length 70 cm. Parallel to the axis of the cylinder, a hole with diameter 0.64 cm is drilled in a radial distance 4.5 cm from the axis of the phantom. A line source insert is also made of polyethylene and it is a tube with the inside diameter 0.32 cm and outside diameter 0.48 cm. The tube may represent known activity and be placed inside the hole of the phantom.

In presented studies, two methods of estimating SF and NECR were used. The first one is based on the ‘true’ Monte Carlo information about the photons’ propagation that is saved in the GATE output file. Having this information it is possible to judge if a coincidence is true, scattered or accidental.

On the other hand, in real measurements not all information about the photons’ propagation is available. Because of that, we apply also a second method as proposed in the NEMA norm, which is based on the analysis of sinograms.



2.5.3. Sensitivity

The sensitivity of a positron emission tomograph is expressed as the rate of true coincidence events T normalized to the total activity A of the source. In order to calculate sensitivity, a linear 1 MBq source of back-to-back gamma photons with length of 70 cm was simulated along the axis of the scanner in the centre of the AFOV.

2.5.4. Image quality

In order to estimate the IQ, simulation of the IEC phantom with cold and hot spheres was performed with the 140 cm long AFOV detector. The phantom was positioned in the center of the AFOV. IQ was calculated based on regions of interest (ROIs) located in two cold and four hot spheres described in NEMA norms (NEMA 2018). Hot spheres of 10 mm (Sphere 10), 13 mm (Sphere 13), 17 mm (Sphere 17) and 22 mm (Sphere 22) diameter, and cold spheres of 28 and 37 mm diameter were simulated. The hot spheres were filled with the activity concentration of 4 times higher than the background region. All spheres centres were positioned in the same transaxial plane located 70 mm from the phantom lid. The 180 mm long cylinder of 51 mm diameter was positioned in the center axis of the phantom. Injection of the 53 MBq of (18)F-FDG dissolved in water was simulated. The scan time was set to 500 s.

In post-processing, positions of interaction points in scintillators were smeared with the FWHM of 5 mm (along axis of the detecting chamber), while the TOF resolution was set to 135 ps (FWHM). All the simulations were performed with GATE (Sarrut *et al* 2014, 2021).

The field of view of the scanner was set to $50 \times 50 \times 130 \text{ cm}^3$. The voxel size was 2.5 mm^3 isotropic. Only true coincidences were taken into reconstruction and their number was about 219 mln. The images were reconstructed with the LM-TOF-MLEM algorithm with the CASTOR (version 3.1) toolkit (Thibaut *et al* 2018). Twenty iterations were used. In the reconstruction, sensitivity and attenuation corrections were included.

Two IQ evaluation metrics were calculated: contrast recovery coefficient (CRC) and background variability (BV). In order to calculate aforementioned parameters, circular ROI was defined on each hot sphere. Furthermore, 12 circular ROIs were also defined on the phantom background. Then, they were replicated to four transaxial slices $\pm 1 \text{ cm}$ and $\pm 2 \text{ cm}$ resulted in 60 ROIs defined in the phantom background for each sphere in total. The CRC for each hot sphere with diameter d was calculated as follows:

$$CRC = \frac{C_{H,d}/C_{B,d} - 1}{4 - 1}, \quad (3)$$




Table 2. Spatial resolution, expressed as FWHM and FWTM of PSF, determined for AFOV of 140 and 200 cm assuming an ideal case that DOI is known.

AFOV	Source position		FWHM (mm)			FWHM (mm)			Voxel dimensions (mm)	
140 cm	0	1	3.8	3.4	4.8	6.3	5.5	10.1	1.56	2.59
		10	3.5	3.4	4.5	5.9	5.6	8.6		
		20	3.8	3.6	4.6	6.1	5.9	8.8		
	52.5 (3/8 AFOV)	1	4.2	3.8	5.8	6.7	6.3	9.3		
		10	3.8	3.7	5.5	6.5	6.4	8.3		
		20	4.0	4.0	5.5	6.5	6.6	8.5		
200 cm	0	1	3.7	3.3	4.9	6.1	5.4	10.2	1.56	2.59
		10	3.4	3.2	4.5	5.7	5.4	8.7		
		20	3.7	3.4	4.7	6.0	5.7	9.0		
	7.5 (3/8 AFOV)	1	4.2	3.7	5.9	6.8	6.1	9.7		
		10	3.8	3.7	5.6	6.4	6.3	8.4		
		20	4.0	3.9	5.6	6.4	6.5	8.6		

where $C_{H,d}$ was the mean counts in the hot sphere and $C_{B,d}$ was the mean of the background ROI counts and 4 in delimiter stands for the true activity ratio between hot spheres and phantom background. The BV for each sphere with diameter d was calculated as follows:

$$BV = \frac{S_d}{C_{B,d}}, \quad (4)$$

where S_d was the standard deviation of the background ROI counts.

2.6. Extension of NEMA norm for TB-PET scanners

As the NEMA-NU-2-2018 norm (and its earlier versions) was provided for classical PET scanners with AFOV less than 50 cm, the sources and phantoms for SF, NECR and sensitivity were limited to 70 cm. However, for total-body PET scanners such sources and phantoms may be not sufficient. Because of that, in discussion

Table 3. Spatial resolution, expressed as FWHM and FWTM of PSF, determined for AFOV of 140 and 200 cm assuming that DOI is not known. For the case of AFOV = 200 cm results for the reconstruction with the limited axial distance between hits ΔZ_{\max} are also presented.

AFOV	Source position		FWHM (mm)			FWHM (mm)			Voxel dimensions (mm)	
140 cm		1	4.9	4.8	7.1	8.2	8.1	19.3	1.56	2.59
		10	6.2	4.8	7.2	10.5	7.9	18.4		
		20	10.3	5.0	7.6	15.3	8.3	18.0		
	52.5 (3/8 AFOV)	1	4.9	4.9	6.3	8.5	8.5	11.1		
		10	6.0	4.8	6.2	10.3	8.0	10.6		
		20	9.6	5.0	6.7	14.9	8.4	12.5		
200 cm		1	4.8	4.8	7.8	8.3	8.0	23.3	1.56	2.59
		10	6.4	4.8	10.7	7.8	21.1			
		20	10.5	5.0	7.9	15.5	8.2	21.3		
	7.5 (3/8 AFOV)	1	4.9	4.8	6.7	8.6	8.2	13.7		
		10	6.2	4.8	6.8	10.4	8.0	12.8		
		20	9.4	5.0	7.1	14.9	8.5	14.7		
200 cm $\Delta Z_{\max} = 30$ cm		1	5.0	4.9	5.4	8.6	8.4	10.8	1.56	2.59
		10	6.1	4.9	10.3	8.2	10.0			
		20	9.5	5.1	5.5	14.8	8.8	10.3		
	75.0 (3/8 AFOV)	1	4.9	4.9	6.2	8.6	8.4	10.8		
		10	6.0	4.9	6.1	10.1	8.3	10.0		
		20	9.2	5.2	6.2	14.7	9.0	10.5		
200 cm $\Delta Z_{\max} = 110$ cm		1	4.9	4.8	7.2	8.3	8.1	19.2	1.56	2.59
		10	6.3	4.8	10.7	7.8	18.1			
		20	10.4	5.0	7.7	15.5	8.2	19.0		
	75.0 (3/8 AFOV)	1	4.9	4.8	6.7	8.6	8.2	13.8		
		10	6.2	4.8	6.8	10.4	8.0	12.8		
		20	9.4	5.0	7.1	14.9	8.5	14.7		

section, results of S, SF and NECR obtained for sources and phantoms elongated to 140 and 200 cm are presented.

3. Results

3.1. Spatial resolution

In order to calculate SR, TOF FBP reconstruction was utilised as explained in section 2. Example images of point-like sources reconstructed with TOF FBP for the J-PET scanner with AFOV = 140 cm are presented in figures 4 and 5. It is visible that achievable SR is in the order of 5 mm. Axial SR (figure 4) improves with the axial distance from the center of the scanner, while the spatial radial resolution worsens with the distance from the scanner axis (figure 5). The values of FWHM and FWTM are given in tables 2–4.

In table 2 the PSF values are shown for the idealized case of known DOI. The result indicates that in the center of the scanner, even with the relatively large scintillators cross section of 6 mm times 30 mm, FWHM values of PSF are equal to 3.3 mm, 3.7 mm and 4.9 mm for the TB-J-PET scanner, for radial, tangential and axial directions, respectively.

Table 3 presents results obtained assuming that DOI is unknown, and the center of the scintillator in the transversal plane is taken as a position of photon interaction. In this case, in the center of the scanner, the FWHM values of PSF, achievable for the assumed geometry, are equal or better than 5 mm (radial, tangential) for both AFOV = 140 cm and AFOV = 200 cm. In case of axial PSF the FWHM values of about 7.1 and 7.8 mm are obtained.

Table 3 also shows results of SR estimated limiting axial (oblique) angle of LOR. Following (Zhang *et al* 2017) the test was performed restricting maximal axial distance ΔZ_{\max} between each pair of hits to 30 and 110 cm. Obtained results indicate that the axial PSF improves from 7.8 to 7.2 mm with limiting axial acceptance distance ΔZ_{\max} to 110 cm, and it improves further to 5.4 mm for $\Delta Z_{\max} = 30$ cm.

Finally the PSF values for the TB-J-PET were determined assuming that DOI can be reconstructed with the resolution described by the Gaussian function with FWHM of 10 mm (table 4). In principle, the WLS layer can allow for determination of DOI (Moskal and Smyrski 2018), yet this solution was not yet tested experimentally. The result presented in table 4 shows an improvement of radial, tangential and axial values of PSF with respect to





Table 4. Spatial resolution, expressed as FWHM and FWTM of PSF, determined for AFOV of 140 and 200 cm assuming that DOI is determined with the resolution of $\text{FWHM}(\text{DOI}) = 10$ mm.

AFOV	Source position		FWHM (mm)			FWHM (mm)			Voxel dimensions (mm)	
140 cm	0	1	4.3	3.8	7.2	6.8	6.2	18.8	1.56	2.59
		10	4.8	3.8	6.6	8.1	6.2	15.4		
		20	6.6	3.8	6.3	12.0	6.2	13.7		
	52.5 (3/8 AFOV)	1	4.3	3.9	6.2	6.9	6.5	10.9		
		10	4.8	3.9	6.0	8.3	6.4	9.8		
		20	7.0	3.8	6.2	12.5	6.3	10.6		
200 cm	0	1	4.3	3.8	7.6	6.9	6.2	22.5	1.56	2.59
		10	4.8	3.8	7.1	8.2	6.2	17.8		
		20	6.6	3.8	6.7	12.0	6.2	16.0		
	75.0 (3/8 AFOV)	1	4.3	3.8	6.6	7.0	6.4	13.3		
		10	4.8	3.9	6.4	8.3	6.4	11.4		
		20	6.8	3.8	6.5	12.3	6.4	12.2		

the case when the DOI is unknown. Specifically the tangential SR is improved, and it is lower than 4 mm for all positions defined in the NEMA norm.

3.2. SF and NECR

As it was described in section 2.5.2 both SF and NECR were estimated using two different methods. In the case of method based on sinograms, the detecting chamber is divided into 1 cm slices and a sinogram is calculated for each slice. After alignment and processing (see figure 6) the number of true coincidences (T) was disentangled from scattered and random (S + R) according to the prescription given in the NEMA norm (see lower panel of figure 6).



To calculate SF and NECR based on the information available from the Monte Carlo simulations, firstly the rates of subsequent types of coincidences were calculated. These rates for *true*, *detector-scattered*, *phantom-scattered* and *accidental* coincidences are shown in figure 7 as a function of the activity concentration, calculated as the total source activity divided by the phantom volume. The determined rates were used to calculate the SF and the NECR as a function of the activity concentration. Results for SF and NECR are presented in table 5 and figure 8, respectively. The SF amounts to about 36% when using a standard length (recommended by NEMA-NU-2018 norm) of the phantom and source. There is no significant difference between obtained SF values for two studied lengths of the detector. Yet, the differences between SF values calculated with two methods are significant. It is due to the differences in the data selection. Narrower data cut in case of sinograms limits the number of false coincidences taken into account. A similar effect, namely that the SF obtained when using sinograms is smaller than the one obtained using *true Monte Carlo* method, was also observed in Yang and Peng (2015).

The NECR characteristics are presented in figure 8. Like in case of SF, the method based on the analysis of sinograms provides different values for NECR than obtained when using a *true Monte Carlo* method. The result obtained according to the methodology required by the NEMA-NU2-2018 norm (NEMA 2018) is shown in the







Table 5. Scatter fraction determined for TB-JPET with AFOV of 140 and 200 cm. Results obtained based on sinograms and on *true Monte Carlo* are presented. Table includes estimations done for source and phantom length (SL) of 70, 140 and 200 cm.

Method	AFOV = 140 cm	AFOV = 200 cm
Sinograms, SL = 70 cm	35.6%	36.2%
True MC, SL = 70 cm	54.2%	54.3%
Sinograms, SL = 140 cm	37.4%	37.6%
True MC, SL = 140 cm	58.4%	58.2%
Sinograms, SL = 200 cm	38.2%	38.0%
True MC, SL = 200 cm	60.0%	59.9%



right panel of figure 8. The maximum value of NECR for a standard length of the phantom is equal to 630 kcps (550 kcps) and is achieved at about 30 kBq cc^{-1} (25 kBq cc^{-1}) for AFOV of 200 cm and 140 cm respectively. The longer the phantom, the lower value of activity concentration for which the NECR peak is obtained.

Table 6. Table summarizing properties of different PET systems. Sources and phantom used for estimating SF, NECR and sensitivity have standard length of 70 cm. TB-J-PET is the geometry investigated in this article (double-layer with AFOV = 200 cm and the plastics strips cross-section of 6 mm times 30 mm), while the J-PET is the geometry investigated in previous studies (AFOV = 100 cm, $D = 75$ cm, double-layer with plastic strips with cross sections of 4 mm times 20 mm) described in detail in Kowalski *et al* (2018).

Scanner	AFOV (cm)	SF (%)	NECR peak (kcps @ kBq cc ⁻¹)	S@0 cm (cps kBq ⁻¹)	TB-FOM (cps kBq ⁻¹)	FWHM @ 1 cm [mm]	
						Transversal	Axial
TB-J-PET (DOI = 10 mm)	200	36.2	630 @ 30	38	19	4.3	7.6
TB-J-PET (DOI not known)						4.8	7.8
TB-J-PET (DOI known)						3.7	4.9
J-PET (Kowalski <i>et al</i> 2018)	100	34.7	300 @ 40	14.9	3.7	3	6
GE: Discovery IQ (5 rings) (Reynés-Llompарт <i>et al</i> 2017, GE 2018)	26	36.2	124 @ 9.1	22.8	1.5	4.2	4.2
Siemens: Biograph mCT (Karlberg <i>et al</i> 2016, Ghabrial <i>et al</i> 2018, Siemens 2018)	21.8	33.2	180.3 @ 29	9.7	0.5	4.4	4.4
Philips: Vereos (Miller 2016, Philips 2020)	16.4	31.6	157.6 @ 52.8	22.1	0.9	3.99	3.99
uExplorer (Spencer <i>et al</i> 2020)	194.0	36.3	1524 @ 17.3	174.0	84.4	3.0	2.8
PennPET Explorer (Vandenberghe <i>et al</i> 2020)	70	32	>1200	55	9.6	4	4

3.3. Sensitivity

Determined values of sensitivities are summarized in figure 9 showing the sensitivity profiles grouped according to the length of the detector. The value of sensitivity determined at the centre of the tomograph amounts to 38 cps kBq⁻¹ and 32 cps kBq⁻¹ for AFOV of 200 cm and 140 cm, respectively. Results for AFOV of 140 cm and SL of 200 cm were omitted.

3.4. Image quality

In figure 10 exemplary reconstructed NEMA IEC images are presented for 5th iteration of TOF-MLEM in axial (left), coronal (middle) and sagittal (right) view. All the spheres could be easily distinguished from the background.

CRC versus BV plots for hot spheres are presented in figure 11. All cases shown, typical trade-off between the CRC and BV parameters. The greatest value of CRC is observed for the biggest sphere (Sphere 22) and the lowest for Sphere 10. Values of the BV started from 0.02 and increased with the number of iterations.

4. Discussion

The performance characteristics presented in the above section were determined applying the prescription given in the NEMA-NU2-2018 norm (NEMA 2018). This norm requires estimations of SF and NECR using 70 cm long linear source and the scatter phantom of the same length. Such length was adjusted formerly to the short AFOV PET scanners being much shorter than 70 cm. In this article we considered total-body PET scanners with the AFOV of 140 and 200 cm and therefore, in addition to NEMA standards we performed estimations of the performance characteristics also for the source and phantom length of 140 and 200 cm. Table 5 indicates that while increasing the length of the phantom from 70 cm to 140 cm and 200 cm, the SF increases only slightly from 35.6% (36.2%) to 37.4% (37.6%) and 38.2% (38%), respectively for AFOV of 140 cm (200 cm). In turn, the maximum of NECR value (figure 8) is decreasing from 550 kcps (630 kcps) to 420 kcps (540 kcps) and 320 kcps (450 kcps) when increasing the phantom and source lengths from 70 cm to 140 cm and 200 cm, respectively for AFOV of 140 cm (200 cm).

The SR was determined for the ideal case when the DOI is known, for the cases when it is reconstructed with the precision of FWHM(DOI) = 10 mm and when it is unknown. In general we observed that in the center of the considered TB-J-PET scanner the radial and tangential resolution is better than 4.9 mm independent of the scenario and that the axial resolution varies between PSF = 4.8 mm (for the ideal case) and PSF = 7.8 mm (for the unknown DOI). In the case when DOI is determined with the uncertainty of 10 mm the radial SR varies between 4.3 mm (at the center) to 6.3 mm at the most distant NEMA defined point. The tangential resolution is equal to about 3.9 mm for all positions and axial PSF varies between 6.4 and 7.6 mm.

In table 6 the characteristics simulated for the double layer TB-J-PET are compared with the state-of-the-art PET scanners and the first TB-PET systems.

The SR estimated for the TB-J-PET (built from strips with 6 mm times 30 mm cross section) is equal to 4.8 mm transversal and 7.6 mm axial. Obtained results show that the SR depends significantly on the DOI resolution and can be improved down to $\text{PSF} = 3.7$ mm (transversal) and 4.8 mm (axial) when improving the DOI reconstruction precision. Thus the values of the SR expected for the TB-J-PET are comparable with the resolution of current clinical PET scanners.

The value of SF of 36.2% obtained for the TB-J-PET is in the range of typical SFs for other systems. The NECR is by 3.5–5 times larger with respect to standard PET and 2.4 times lower than of uExplorer.

The sensitivity in the center of the scanner is higher by 1.7–3.9 times with respect to current PET systems and lower by a factor of 4.6 and 1.4 with respect to uExplorer and PennPET Explorer, respectively. Yet, the sensitivity of TB-J-PET may be further increased by adding a third detection layer.

Reconstructed images of the NEMA IEC phantom and the CRC versus BV curves shows that good quality images could be obtained with the total body J-PET setup. However, the impact of other effects (point spread modelling, post-reconstruction filtering, different reconstruction algorithms, other corrections etc) have to be further carefully investigated.

In case of the total-body scan, the figure of merit (TB-FOM), a measure of the whole-body sensitivity, can be expressed as the rate of events registered and selected for the image reconstruction per rate of photons emitted from the whole-body. Taking into account an approximately triangular shape of the sensitivity profile, in the case of the 200 cm long source TB-FOM can be approximated as $\text{TB-FOM} = 0.5$ times $S(@0\text{cm})$ times $\text{AFOV} / 200$ cm. Comparing values of TB-FOM shown in table 6 one can infer that the sensitivity for the total-body scan (TB-FOM) of the J-PET is by factor of 12.6–8 larger with respect to TB-FOM of the current clinical PET scanners (in agreement with estimations presented in articles (Moskal and Stępień 2020, Vandenberghe *et al* 2020)), and it is by factor of 4.4 less than TB-FOM of uExplorer.

5. Conclusions

Performance characteristics of the total-body J-PET scanner (TB-J-PET) built from plastic scintillators were estimated according to the NEMA NU 2-2018 standards (NEMA 2018). The calculations were performed by means of the GATE simulation package (Jan *et al* 2004, 2011, Sarrut *et al* 2014, 2021) and the J-PET software analysis tools (Kowalski *et al* 2018, Krzemień *et al* 2020). For the TB-J-PET a double layer geometry with plastic strips with dimensions of 6 mm times 30 mm was assumed. The strips are arranged in modules forming 24-sided polygon with the inner diameter of 78.6 cm.

The performed simulations indicated that for the TB-J-PET with $\text{AFOV} = 200$ cm a SR of $\text{PSF} = 3.7$ mm (transversal) and $\text{PSF} = 4.9$ mm (axial) is achievable. The NECR peak of 630 kcps is expected at 30 kBq cc^{-1} activity concentration, and the sensitivity at the center amounts to 38 (cps kBq^{-1}). The SF is estimated to 36.2%.

The values of SF and SR are comparable to those obtained for the clinical PET scanners as well as to the first total-body uExplorer PET. However, the NECR is by factor of about 4 larger with respect to standard PET systems. Moreover, the TB-J-PET sensitivity for the whole-body scan (TB-FOM) is by a factor of 12.6–8 larger with respect to current clinical PET systems with AFOV in the range from 16 cm to 26 cm, and by a factor of about 4.9 less than of uExplorer PET.

Though the whole-body sensitivity increase of the TB-J-PET is less than the one of uExplorer it is still a significant improvement with respect to the current 16–26 cm long PET systems (and can still be increased by adding a third detection layer). This makes J-PET an economic and alternative technology for the construction of the total-body PET systems. The cost of the TB-J-PET is expected to be a factor of 5 lower with respect to the LYSO based TB-PET systems (Moskal and Stępień 2020). This is mainly due to the application of the axially arranged long strips of plastic scintillators with readout at the edges, instead of the detectors built from radially arranged blocks of heavy scintillator crystals, thus reducing significantly not only the cost of scintillators but also the number of SiPMs and electronics channels. In addition to the discussed advantages, the mechanical robustness of plastics with respect to crystals allows for making the plastic total-body scanner lightweight, modular and portable, thus making TB-J-PET a promising cost-effective solution for the broad clinical applications of total-body PET scanners.

Sensitivity, SF and NECR were obtained not only for standard NEMA sources and phantoms with 70 cm length but also for longer ones with lengths of 140 and 200 cm equal to the length of scanners investigated in this article. We propose that the NECR characteristic should be redefined and the characteristic should be presented not for activity concentration but activity concentration normalized by the length of the phantom or just for the total activity used.

Acknowledgments

The authors acknowledge technical and administrative support of A. Heczko, M. Kajetanowicz and W. Migdał. This work was supported by the National Centre for Research and Development through grant INNOTECH-K1/IN1/64/159174/NCBR/12, the Foundation for Polish Science through the TEAM POIR.04.04.00-00-4204/17 programmes, the National Science Centre of Poland through grants no. 2017/25/N/NZ1/00861, 2019/35/B/ST2/03562, the Ministry of Science and Higher Education through grants no. 6673/IA/SP/2016, 7150/E-338/SPUB/2017/1, the Austrian Science Fund through grant no. FWF-P26783 and the Jagiellonian University via project CRP/0641.221.2020. JB acknowledges the support of InterDokMed project no. POWR.03.02.00-00-I013/16.

Disclosure of conflicts of interest

The authors have no relevant conflicts of interest to disclose.

ORCID iDs

P Moskal  <https://orcid.org/0000-0002-4229-3548>

P Kowalski  <https://orcid.org/0000-0002-3545-3865>

E Czerwiński  <https://orcid.org/0000-0002-5076-3175>

B C Hiesmayr  <https://orcid.org/0000-0001-9062-6039>

E Stępień  <https://orcid.org/0000-0003-3589-1715>

References

- Kishimoto A et al 2017 First demonstration of multi-color 3-D *in vivo* imaging using ultra-compact Compton camera *Sci. Rep.* **7** 2045–322
- Badawi R D et al 2019 First human imaging studies with the EXPLORER total-body PET *J. Nucl. Med.* **60** 299–303
- Bailey D L, Townsend D W, Valk P E and Maisey M N 2005 *Positron Emission Tomography, Basic Sciences* (Berlin: Springer)
- Becker R et al 2017 The SAFIR experiment: concept, status and perspectives *Nucl. Instrum. Methods Phys. Res. A* **845** 648–51
- Borja A J et al 2020 An update on the role of total-body PET imaging in the evaluation of atherosclerosis *PET Clin.* **15** 477–85
- Brunner S E and Schaart D R 2017 BGO as a hybrid scintillator/Cherenkov radiator for cost-effective time-of-flight PET *Phys. Med. Biol.* **62** 4421–39
- Cates J W and Levin C S 2019 Electronics method to advance the coincidence time resolution with bismuth germanate *Phys. Med. Biol.* **64** 175016
- Chen J, Zhang L and Zhu R-Y 2004 Large size LYSO crystals for future high energy physics experiments *IEEE Symposium Conference Record Nuclear Science 2004 1* (Piscataway, NJ: IEEE) 117–125
- Cherry S R et al 2017 Total-body imaging: transforming the role of positron emission tomography *Sci. Transl. Med.* **9** 1946–6234
- Cherry S R et al 2018 Maximizing sensitivity to create new opportunities for clinical research and patient care *J. Nucl. Med.* **59** 3–12
- Conti M 2009 State of the art and challenges of time-of-flight pet *Phys. Med.* **25** 1–11
- Conti M et al 2005 First experimental results of time-of-flight reconstruction on an LSO PET scanner *Phys. Med. Biol.* **50** 4507–4526
- Crystals S-G 2016 BGO Scintillation Material Data Sheet
- Crystals S-G 2018a LYSO Scintillation Material Data Sheet
- Crystals S-G 2018b BC-400, BC-404, BC-408, BC-412, BC-416 Plastic Scintillators Data Sheet
- Crystals S-G <https://crystals.saint-gobain.com>
- Donnard J et al 2012 Compton imaging with liquid xenon and ⁴⁴Sc: recent progress toward 3 gamma imaging *Nucl. Med. Rev.* **15** C64–C67
- Efthimiou N 2020 New challenges for PET image reconstruction for total-body imaging *PET Clin.* **15** 453–461
- Eljen Technology Physical constants of plastic scintillators (https://eljentechnology.com/images/technical_library/Physical_Constants_Plastic.pdf)
- Gajos A 2020 Sensitivity of discrete symmetry tests in the positronium system with the J-PET detector *Symmetry* **12** 1268–1282
- GE. Producer's website with technical details (http://gehealthcare.com/en/products/categories/molecular_imaging/pet-ct/pet-ct_scanners/discovery_iq) 2018-03-29
- Ghabrial A, Franklin D and Zaidi H 2018 A Monte Carlo simulation study of the impact of novel scintillation crystals on performance characteristics of PET scanners *Phys. Med.* **50** 37–45
- Gonzalez-Montoro A et al 2017 Highly improved operation of monolithic BGO-PET blocks *J. Instrum.* **12** C11027
- Grant A M et al 2016 NEMA NU 2-2012 performance studies for the SiPM-based TOF-PET component of the GE SIGNA PET/MR system *Med. Phys.* **43** 2334–2343
- Grignon C et al 2007 Nuclear medical imaging using β^+ γ coincidences from ⁴⁴Sc radio-nuclide with liquid xenon as detection medium *Nucl. Instrum. Methods Phys. Res. A* **571** 142–5
- Gundacker S et al 2020 Experimental time resolution limits of modern SiPMs and TOF-PET detectors exploring different scintillators and Cherenkov emission *Phys. Med. Biol.* **65** 025001
- Hamidreza H, Alireza K-A, Mohammadreza A and Pardis G 2017 Compton scatter tomography in ToF-PET *Phys. Med. Biol.* **62** 7641
- Hiesmayr B C and Moskal P 2019 Witnessing entanglement in Compton scattering processes via mutually unbiased bases *Sci. Rep.* **9** 1–14
- Houshmand S et al 2015 An update on novel quantitative techniques in the context of evolving whole-body pet imaging *PET Clin.* **10** 45–58
- Jan S et al 2004 GATE: a simulation toolkit for PET and SPECT *Phys. Med. Biol.* **49** 4543–61
- Jan S et al 2011 GATE v6: a major enhancement of the GATE simulation platform enabling modelling of CT and radiotherapy *Phys. Med. Biol.* **56** 881–901

- Jones T 2020 Total body pet imaging from mice to humans *Frontiers Phys.* **8** 77
- Kamińska D et al 2016 A feasibility study of ortho-positronium decays measurement with the J-PET scanner based on plastic scintillators *Eur. Phys. J. C* **76** 1–14
- Karlberg A M et al 2016 Quantitative comparison of PET performance - Siemens Biograph mCT and mMR *Eur. J. Nucl. Med. Mol. Imaging Phys.* **3** 5
- Karp J S et al 2020 Pennpet explorer: design and preliminary performance of a whole-body imager *J. Nucl. Med.* **61** 136–43
- Kenji S, Masao Y, Yusuke O, Mizuki U, Kenichiro O, Tadashi O, Hiroyuki T, Kei K, Akira Y and Miwako T 2020 Development of simultaneous pet and compton imaging using gagg-sipm based pixel detectors *Nucl. Instrum. Methods Phys. Res. A* **954** 161499
- Khateri P et al 2019 Implementation of cylindrical pet scanners with block detector geometry in stir *EJNMMI Phys.* **6** 15
- Kowalski P et al 2015 Multiple scattering and accidental coincidences in the j-pet detector simulated using gate package *Acta Phys. Pol. A* **127** 1505–12
- Kowalski P et al 2016 Scatter fraction of the j-pet tomography scanner *Acta Phys. Pol.* **B47** 549
- Kowalski P et al 2018 Estimating the NEMA characteristics of the J-PET tomograph using the GATE package *Phys. Med. Biol.* **63** 165008
- Krzemien W et al 2020 J-PET Framework: Software platform for PET tomography data reconstruction and analysis *SoftwareX* **11** 100487
- Kuramoto M, Nakamori T, Kimura S, Gunji S, Takakura M and Kataoka J 2017 Development of tof-pet using compton scattering by plastic scintillators *Nucl. Instrum. Methods Phys. Res. A* **845** 668–72
- Kwon S I et al 2019 Dual-ended readout of bismuth germanate to improve timing resolution in time-of-flight pet *Phys. Med. Biol.* **64** 105007
- Lang C, Habs D, Parodi K and Thirolf P G 2014 Sub-millimeter nuclear medical imaging with high sensitivity in positron emission tomography using β^+ γ coincidences *J. Instrum.* **9** P01008
- Lang C, Habs D, Thirolf P G and Zoglauer A 2012 Submillimeter nuclear medical imaging with a Compton Camera using triple coincidences of collinear/beta+ annihilation photons and/gamma-rays arXiv:1202.0397
- Majewski S 2020 Imaging is believing: the future of human total-body molecular imaging starts now *Il Nuovo Cimento* **100** 43–58
- Mao R, Zhang L and Zhu R-Y 2008 Optical and scintillation properties of inorganic scintillators in high energy physics *IEEE Trans. Nucl. Sci.* **55** 2425–31
- Mao R et al 2013 Crystal growth and scintillation properties of lso and lyso crystals *J. Cryst. Growth* **368** 97–100
- McKenney-Drake M L et al 2018 18 f-naf and 18 f-fdg as molecular probes in the evaluation of atherosclerosis *Eur. J. Nucl. Med. Mol. Imaging* **45** 2190–200
- Miller M A 2016 Focusing on high performance *Philips Brochure* (www.philips.com/VereosPETCT)
- Mizuki U et al 2020 Double photon emission coincidence imaging with GAGG-SIPM Compton camera *Nucl. Instrum. Methods Phys. Res. A* **954** 161682
- Moskal P and Stępień E Ł 2020 Prospects and clinical perspectives of total-body PET imaging using plastic scintillator *PET Clin.* **15** 439–52
- Moskal P, Jasińska B, Stępień E and Bass S D 2019a Positronium in medicine and biology *Nat. Rev. Phys.* **1** 527–9
- Moskal P and Smyrski J 2018 Detecting device for determining a position of reaction of gamma quanta and a method for determining a position of reaction of a gamma quanta in positron emission tomography *US Patent* 10,042,058
- Moskal P et al 2011 Novel detector systems for the positron emission tomography *Bio-Algorithms Med-Syst.* **7** 73–8
- Moskal P et al 2014 Test of a single module of the J-PET scanner based on plastic scintillators *Nucl. Instrum. Methods A* **764** 317–21
- Moskal P et al 2015 A novel method for the line-of-response and time-of-flight reconstruction in TOF-PET detectors based on a library of synchronized model signals *Nucl. Instrum. Methods A* **775** 54–62
- Moskal P et al 2016a Time resolution of the plastic scintillator strips with matrix photomultiplier readout for J-PET tomograph *Phys. Med. Biol.* **61** 2025–47
- Moskal P et al 2016b Potential of the J-PET detector for studies of discrete symmetries in decays of positronium atom-a purely leptonic system *Acta Phys. Pol. B* **47** 509–35
- Moskal P et al 2018 Feasibility studies of the polarization of photons beyond the optical wavelength regime with the J-PET detector *Eur. Phys. J. C* **78** 970–8
- Moskal P et al 2019b Feasibility study of the positronium imaging with the J-PET tomograph *Phys. Med. Biol.* **64** 055017
- Moskal P et al 2020 Performance assessment of the 2 γ positronium imaging with the total-body pet scanners *EJNMMI Phys.* **7** 44
- Nakajima R, Abe K and Sakai S 2017 Igg4-related diseases; whole-body fdg-pet/ct may be easier to evaluate rare lesions *J. Nucl. Med.* **58** 943–943
- Nardo L et al 2020 Potential roles of total-body pet/computed tomography in pediatric imaging *PET Clin.* **15** 271–9
- National Institute of Standards and Technology 2020 (<https://nist.gov/pml>)
- NEMA 2018 Nema standards publication nu 2-2018: Performance measurements of positron emission tomographs (pet) *Natl Electr. Manuf. Assoc.* (NEMA NU 2-2018)
- Niedźwiecki S et al 2017 J-PET: a new technology for the whole-body PET imaging *Acta Phys. Pol. B* **48** 1567
- Oger T et al 2012 A liquid xenon TPC for a medical imaging Compton telescope *Nucl. Instrum. Methods Phys. Res. A* **695** 125–8
- Philips. Philips Vereos PET/CT brochure (<https://philipsproductcontent.blob.core.windows.net/assets/20170523/360753349c5d4a6aa46ba77c015e75b4.pdf>) 2020-06-29
- Reynés-Llompant G et al 2017 Performance characteristics of the whole-body Discovery IQ PET/CT system *J. Nucl. Med.* **58** 1155–61
- Sarrut D et al 2014 A review of the use and potential of the GATE Monte Carlo simulation code for radiation therapy and dosimetry applications *Med. Phys.* **41** 064301
- Sarrut D et al 2021 Advanced Monte Carlo simulations of emission tomography imaging systems with GATE *Phys. Med. Biol.* **66** 10TR03
- Schmall J P, Karp J S and Alavi A 2019 The potential role of total body pet imaging in assessment of atherosclerosis *PET Clin.* **14** 245–50
- Sharma N G et al 2020a Hit-time and hit-position reconstruction in strips of plastic scintillators using multi-threshold readouts *IEEE Trans. Radiat. Plasma Med. Sci.* **4** 528
- Sharma S et al 2020b Estimating relationship between the Time Over Threshold and energy loss by photons in plastic scintillators used in the J-PET scanner *EJNMMI Phys.* **7** 1–15
- Shivani S, Łuczynska E, Heinze S and Moskal P 2020 Development of j-pem for breast cancer detection *Acta Phys. Pol. A* **137**
- Shopa R Y 2020 Estimation of spatial resolution for 3-layer j-pet scanner using tof fbp based on event-by-event approach *Acta Phys. Pol. B* **51**
- Shopa R Y et al 2017 Three-dimensional image reconstruction in j-pet using filtered back-projection method *Acta Phys. Polon. B* **48** 1757
- Siemens. Online catalogue with technical details (http://activexray.com/pdf/Siemens_Biograph.pdf) 2018-03-29
- Smyrski J et al 2014 Application of wls strips for position determination in strip pet tomograph based on plastic scintillators *Bio-Algorithms Med-Syst.* **10** 59–63
- Smyrski J et al 2017 Measurement of gamma quantum interaction point in plastic scintillator with WLS strips *Nucl. Instrum. Methods A* **851** 39–42

- Spencer B A *et al* 2020 Performance evaluation of the uEXPLORER Total-body PET/CT scanner based on NEMA NU 2-2018 with additional tests to characterize long axial field-of-view PET scanners *J. Nucl. Med.* **120** 120
- Surti S, Pantel A R and Karp J S 2020a Total body pet: Why, how, what for? *IEEE Trans. Radiat. Plasma Med. Sci.* **4** 283–92
- Surti S, Werner M E and Karp J S 2013 Study of pet scanner designs using clinical metrics to optimize the scanner axial fov and crystal thickness *Phys. Med. Biol.* **58** 3995
- Surti S *et al* 2020b Benefit of improved performance with state-of-the art digital pet/ct for lesion detection in oncology *J. Nucl. Med.* **120** jnumed
- Thibaut M, Simon S, Didier B, Julien B, Thomas C, Claude C, Marina F, Frédéric L and Dimitris V 2018 Castor: a generic data organization and processing code framework for multi-modal and multi-dimensional tomographic reconstruction *Phys. Med. Biol.* **63** 185005
- Thielemans K *et al* 2012 STIR: software for tomographic image reconstruction release 2 *Phys. Med. Biol.* **57** 867
- Thirolf P G, Lang C and Parodi K 2015 Perspectives for highly-sensitive pet-based medical imaging using β^+ γ coincidences *Acta Phys. Pol. A* **127**
- Van Sluis J *et al* 2019 Performance characteristics of the digital biograph vision pet/ct system *J. Nucl. Med.* **60** 1031–6
- Vandenberghe S, Moskal P and Karp J S 2020 State of the art in total body pet *EJNMMI Phys.* **7** 35
- Vilardi I *et al* 2006 Optimization of the effective light attenuation length of yap: Ce and lyso: Ce crystals for a novel geometrical pet concept *Nucl. Instrum. Methods Phys. Res. A* **564** 506–14
- Yamashita H, Kubota K and Mimori A 2014 Clinical value of whole-body pet/ct in patients with active rheumatic diseases *Arthritis Res. Ther.* **16** 423
- Yang X and Peng H 2015 The use of noise equivalent count rate and the NEMA phantom for PET image quality evaluation *Phys. Med.* **31** 179–84
- Yoshida E *et al* 2020 Whole gamma imaging: a new concept of PET combined with Compton imaging *Phys. Med. Biol.* **65** 125013
- Zein S A *et al* 2020 Physical performance of a long axial field-of-view pet scanner prototype with sparse rings configuration: A monte carlo simulation study *Med. Phys.* **47** 1949–57
- Zhang J, Knopp M I and Knopp M V 2019 Sparse detector configuration in sipm digital photon counting pet: a feasibility study *Mol. Imaging Biol.* **21** 447–53
- Zhang X *et al* 2017 Quantitative image reconstruction for total-body PET imaging using the 2-meter long EXPLORER scanner *Phys. Med. Biol.* **62** 2465
- Zhang X *et al* 2019 Total-body parametric imaging using kernel and direct reconstruction on the uexplorer *J. Nucl. Med.* **60** 456–456
- Zhang X *et al* 2020 Total-body dynamic reconstruction and parametric imaging on the uexplorer *J. Nucl. Med.* **61** 285–91
- Zhang Y and Wong W-H 2017 System Design Studies for a Low-cost High-resolution BGO PET with 1-meter Axial Field of View *J. Nucl. Med.* **58** 221–221